

Compliance with confidence







0764 Cert. No. 97 ISO 9001

## **SPECIFICATION (S0133)**

Pen Type thermometer

**Product Code ACTX1010** 

## **Product Features**

- Suitable for use in Food Processing, Laboratory, Brewing etc.
- Fast temperature response
- Stainless steel probe
- Splash proof



## **Technical Data**

Measurement range	-50 to 200°C			
Accuracy	±1°C between -50 to 150°C otherwise ±2°C			
Display resolution	0.1° between -19.9°C to 199.9°C otherwise 1°			
Sampling rate	Every 0.7 second			
Temperature response	Within 8 seconds from 25°C raised to 100°C			
Automatic shut off	After idle for 10 minutes			
Battery & Life	1.5V LR44 x 1. Continuous operation 5000 hours			
Ambient temperature	-10 to 50°C			
Unit dimensions	Ø22 x 77mm			

## Operation

Measurement

- 1. Press the hold button to switch thermometer on
- 2. Probe into test object about 25mm depth
- 3. Take the reading when stable
- 4. Press and hold the hold button for 4 seconds and release to switch off the thermometer.

Data hold function

- 1. Press the hold button once to retain reading, degree symbol flashes.
- 2. Press hold button again to release the HOLD function.

Maximum and minimum temperature memory

- 1. Press min/max button once to display max reading, MAX symbol flashes.
- 2. Press min/max button again to display min reading. MIN symbol flashes.
- 3. Press min/max button once back to normal temperature display.
- 4. Turn thermometer off to clear memory.

This document was prepared on behalf of Klipspringer Ltd and the information included is to the best of our knowledge correct at the time of writing. Klipspringer offers the information within this document as a guide only, they do not represent any guarantee of the prescribed products in the sense of the legal guarantee regulations. It is the responsibility of the end user to ensure the items purchased are suitable for the intended application.

Supplier/Importer: Klipspringer Ltd

Address: Rynor House, Farthing Road, Ipswich, Suffolk, UK. IP1 5AP

Telephone: +44 (0)1473 461800

Date of issue	16/10/2019	Revision No.	001	Revised by	SB
Authorised by	G. Bush	Document No.	S0133	Page 1 of 1	