

Compliance with confidence







Cert. No

Product Specification: pH Electrode (S059)

P14 General Purpose pH Electrode

Broad spectrum of pH testing including food and beverages such as fruit juices and liquids

PH Range 0-14
Cable Length 1 Metre

E O Point -25mV to +25mV

Connector BNC (others available upon request)

Temp Range °C 0-80°C Combination Yes

Response Time <30 secs 95% FSD

Body Type Epoxy

Ref. Type AgCl, Gel filled Single Junction Application General Purpose, field use.

Junction Type Porous Teflon
Storage Solution PEB-216
Shaft Length 120mm ± 2mm
Recommended Cleaning Solution PEB215

Stem Dia mm 11.8 – 12.2mm



This document was prepared on behalf of Klipspringer Ltd and the information included is to the best of our knowledge correct at the time of writing. Klipspringer offers the information within this document as a guide only, they do not represent any guarantee of the prescribed products in the sense of the legal guarantee regulations. It is the responsibility of the end user to ensure the items purchased are suitable for the intended application.

Supplier/Importer: Klipspringer Ltd

Address: Rynor House, Farthing Road, Ipswich, Suffolk, UK. IP1 5AP

Telephone: 0044 (0) 1473 461800

Email/Website: info@klipspringer.com / www.klipspringer.com

Sheena Britton

Technical Compliance Manager

Klipspringer Ltd 20-03-17

Date of issue	20-03-17	Revision No.	001	Revised by	SB
Authorised by	S. Britton	Document No.	S059	Page 1 of 1	