



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No 5658
Buffer Solution
pH 10.00 [Borate] BLUE
'BAKER ANALYZED'
Lot No 1732001862
Release Date 21-Nov-2017

Test	Specification	Actual value
pH at 25°C	9.95 - 10.05	9.96

Temperature (°C)

0 10.30 pH

10 10.17 pH

20 10.05 pH

25 10.00 pH

30 9.96 pH

40 9.89 pH

50 9.82 pH

60 9.76 pH

70 9.72 pH

80 9.68 pH

90 9.64 pH

Contains boric acid / potassium chloride / sodium hydroxide

Contains 0.01% sodium benzoate.

This buffer is traceable to NIST Standard Reference Buffer.

Use Before Date: 2019-11

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.


Jamie Ethier
Vice President Global Quality