



Laboratory chemicals and products-Clinical chemistry and hematology products-Fine & industrial chemicals

Certificate of Analysis

Product No 5647
Buffer Solution
pH 7.00 [Phosphate] GREEN
'BAKER ANALYZED'

Lot No 1708901865
Release Date 06-Apr-2017

Test	Specification	Actual value
pH at 20°C	6.98 - 7.02	7.00

Temperature (°C)

0 7.13 pH
10 7.05 pH
20 7.00 pH
25 6.98 pH
30 6.98 pH
40 6.97 pH
50 6.96 pH
60 6.96 pH
70 6.97 pH
80 6.98 pH
90 7.00 pH

Contains potassiumdihydrogenphosphate /
disodiumhydrogenphosphate

Contains 0.01% sodium benzoate.

This buffer is traceable to NIST Standard Reference Buffer.

Use Before Date: 2019-03

For Laboratory, Research or Manufacturing Use.

This document is automatically generated.

For questions on this cofa, please contact Customer Service at avantor.emea@avantormaterials.com

The information in this document is property of Avantor Performance Materials Poland S.A.

Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel:+48 322392000

Only for Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) relevant countries:

For information on Use Descriptors please contact us by mail to

JTBaker_REACH_USERS@avantormaterials.com indicating the JTBaker product number you purchase.

Jamie Ethier
Vice President Global Quality