

CERTIFICATE OF ANALYSIS

Material Description: Buffer Solution pH 10 Colour Coded (Blue)

CAS Number: Mixture

Batch Number: 899564

| Supplied As: | Product Code | Product Description |
|--------------|--------------|--------------------------------------------------|
| | PEB304-10 | Buffer Solution pH 10 Colour Coded (Blue) 2.5L. |
| | PEB404-10 | Buffer Solution pH 10 Colour Coded (Blue) 1L. |
| | PEB504-10 | Buffer Solution pH 10 Colour Coded (Blue) 500ml. |

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

| Specification | | | | | | Test Results | |
|---------------|-------|---------------|---------------|---------------|-------|--------------|--------|
| Parameter | Units | Minimum value | Maximum value | Typical value | Notes | Result | Status |
| pH @ 20°C | pH | 9.99 | 10.01 | 10.00 | | 9.993 | Pass |

| Additional Information | | | | | |
|------------------------|-------|---------------|---------------|---------------|-----------|
| Parameter | Units | Minimum value | Maximum value | Typical value | Notes |
| pH @ 10°C | pH | | | 10.10 | |
| pH @ 15°C | pH | | | 10.04 | |
| pH @ 25°C | pH | | | 9.94 | |
| pH @ 30°C | pH | | | 9.90 | |
| pH @ 37°C | pH | | | 9.86 | |
| pH @ 40°C | pH | | | 9.82 | |
| pH @ 45°C | pH | | | 9.79 | |
| pH @ 50°C | pH | | | 9.79 | |
| NIST SRM 186-I-g | | | | | Traceable |
| NIST SRM 186-II-g | | | | | Traceable |
| NIST SRM 191d-I | | | | | Traceable |
| NIST SRM 191d-II | | | | | Traceable |

Doc. tracking ref: 2628/CA/899564/001
Issued by: Radek Tameczka
Position: Laboratory Manager
Issue Date: 15th January 2024

Certificate generated electronically and is valid without a signature.

Klipspringer Ltd. Registered in England and Wales: Klipspringer Ltd, Rynor House, Farthing Road, Ipswich, IP1 5AP. Registration number: 07676073.