

## Certificate of Analysis

### Product Information:

<b>Product Code</b>	ECBU1BT
<b>Description</b>	1.68 pH @25 deg C pH Buffer Solution
<b>Batch Number</b>	053/03
<b>Manufacturing Date</b>	03-Feb-2023
<b>Expiration Date</b>	03-Feb-2026

### Actual Analysis

Measurement	Acceptance Range	Results
pH value (@ 25 deg C)	1.68+/-0.01 pH	Pass

### Method of Standardization

A mV instrument with a sensor calibrated to NIST traceable standards was used for cross checking the solution value.

Approved By

**ThermoFisher Scientific**  
**Water and Lab Products**  
**Quality Assurance**

This is a computer-generated document. No signature is required

Rev A