

CERTIFICATE OF ANALYSIS

Material Description: Buffer Solution pH 7 Colour Coded (Yellow)

CAS Number: Mixture

Batch Number: 916814

Supplied As:

Product Code

Product Description

PEB304-07

Buffer Solution pH 4 Colour Coded (Yellow) **2.5l.**

PEB404-07

Buffer Solution pH 4 Colour Coded (Yellow) **1l.**

PEB504-07

Buffer Solution pH 4 Colour Coded (Yellow) **500ml.**

PEB1104-07

Buffer Solution in Dosing Bottle pH 7.0 (Yellow) **500ml.**

PEB1204-07

Buffer Solution in Dosing Bottle pH 7.0 (Yellow) **1l.**

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

Specification						Test Results	
Parameter	Units	Minimum value	Maximum value	Typical value	Notes	Result	Status
pH @ 20°C	pH	6.99	7.01	7.00		6.997	Pass
Additional Information							
Parameter	Units	Minimum value	Maximum value	Typical value	Notes		
pH @ 10°C	pH			7.05			
pH @ 15°C	pH			7.02			
pH @ 25°C	pH			6.99			
pH @ 30°C	pH			6.98			
pH @ 37°C	pH			6.97			
pH @ 40°C	pH			6.97			
pH @ 45°C	pH			6.96			
pH @ 50°C	pH			6.96			
NIST SRM 185i							Traceable
NIST SRM 186-I-g							Traceable
NIST SRM 186-II-g							Traceable
Traceable to NIST							

Doc. tracking ref: 2626/CA/873636/001

Issued by: Radek Tameczka

Position: Laboratory Manager

Issue Date: 10th Feb 2023

Certificate generated electronically and is valid without a signature.