

Compliance with confidence

Klipspringer Ltd Rynor House, Farthing Road Ipswich, IP1 5AP, UK

O1473 461 800

sales@klipspringer.com

www.klipspringer.com

CERTIFICATE OF ANALYSIS

Material Description: Buffer Solution pH 2 Colour Coded (Green)

CAS Number: Mixture **Batch Number:** 866920

Supplied As: Product Code Product Description

> PEB604-02 Buffer Solution pH 2 Colour Coded (Green) 1L.

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

Specification	Test Results						
Parameter	Units	Minimum value	Maximum value	Typical value	Notes	Result	Status
pH @ 20°C	pН	1.99	2.01	2.00		2.002	Pass

Additional Information								
Parameter	Units	Minimum value	Maximum value	Typical value	Notes			
pH @ 10°C	pН			1.98				
pH @ 15°C	pН			2.02				
pH @ 25°C	рН			2.00				
pH @ 30°C	pН			2.00				
pH @ 35°C	рН			2.02				
pH @ 40°C	pН			2.01				
pH @ 45°C	pН			2.02				
pH @ 50°C	рН			2.02				
NIST SRM 189c					Traceable			
NIST SRM 185i					Traceable			

Doc. tracking ref: 0845/CA/866920/001

Issued by: Radek Tameczka Position: Laboratory Manager **Issue Date:** 20th January 2023

Certificate generated electronically and is valid without a signature.

Klipspringer Ltd. Registered in England and Wales: Klipspringer Ltd, Rynor House, Farthing Road, Ipswich, IP1 5AP. Registration number: 07676073.



