

Compliance with confidence

Klipspringer Ltd Rynor House, Farthing Road Ipswich, IP1 5AP, UK

O1473 461 800

sales@klipspringer.com

www.klipspringer.com

CERTIFICATE OF ANALYSIS

Material Description: Buffer Solution pH 4 Colour Coded (Red)

CAS Number: Mixture **Batch Number:** 867293

Product Code **Supplied As: Product Description**

PEB304-04 Buffer Solution Ph 4 Colour Coded (Red) 2.51. Buffer Solution Ph 4 Colour Coded (Red) 11. PEB404-04 PEB504-04 Buffer Solution Ph 4 Colour Coded (Red) 500ml.

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

Specification						Test Res	Test Results	
Parameter	Units	Minimum value	Maximum value	Typical value	Notes	Result	Status	
pH @ 20°C	рН	3.99	4.01	4.00		4.010	Pass	
Additional Info	ormation							
Parameter	Units	Minimur value	n Maximu value	m Typical value	Notes			
pH @ 10°C	рН			4.00				
pH @ 15°C	рН			4.00				
pH @ 25°C	рН			4.00				
pH @ 30°C	рН			4.01				
pH @ 37°C	рН			4.02				
pH @ 40°C	рН			4.03				
pH @ 45°C pH				4.04				
pH @ 50°C	рН			4.05				
NIST SRM 185i					Traceable			
NIST SRM 186-I-g					Traceable			
NIST SRM 186-II-g					Traceable			

Doc. tracking ref: 2626/CA/867293/001 Issued by: Radek Tameczka **Position:** Laboratory Manager **Issue Date:** 10th Feb 2023

Certificate generated electronically and is valid without a signature.

Klipspringer Ltd. Registered in England and Wales: Klipspringer Ltd, Rynor House, Farthing Road, Ipswich, IP1 5AP. Registration number: 07676073.



