

Compliance with confidence

Klipspringer Ltd Rynor House, Farthing Road Ipswich, IP1 5AP, UK

01473 461 800sales@klipspringer.com

www.klipspringer.com

CERTIFICATE OF ANALYSIS

Material Description:Buffer Solution pH 4 Colour Coded (Red)CAS Number:MixtureBatch Number:873636Supplied As:Product Code
PEB304-04PEB404-04
PEB504-04Buffer Solution Ph 4 Colour Coded (Red) 2.5I.
Buffer Solution Ph 4 Colour Coded (Red) 1I.
Buffer Solution Ph 4 Colour Coded (Red) 500ml.

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

Specification						Test Results	
Parameter	Units	Minimum value	Maximum value	Typical value	Notes	Result	Status
оН @ 20°С рН		3.99	4.01	4.00		3.991	Pass
Additional Info	ormation	· · · · ·					
Parameter L		Minimun value	n Maximu value	m Typica value	Notes		
pH @ 10°C	pН			4.00			
pH @ 15°C pH				4.00			
рН @ 25°С рН				4.00			
рН @ 30°С рН				4.01			
рН @ 37°С рН				4.02			
рН @ 40°С рН				4.03			
рН @ 45°С рН				4.04			
рН @ 50°С рН				4.05			
NIST SRM 185i					Traceable		
NIST SRM 186-I-g					Traceable		
NIST SRM 186-II-g					Traceable		
Traceable to NIS	ST						

Doc. tracking ref:2626/CA/873636/001Issued by:Radek TameczkaPosition:Laboratory ManagerIssue Date:10th Feb 2023

Certificate generated electronically and is valid without a signature.

Klipspringer Ltd. Registered in England and Wales: Klipspringer Ltd, Rynor House, Farthing Road, Ipswich, IP1 5AP. Registration number: 07676073.

Shadow Boards & Visual Management Segregation Screens Hygiene & Cleaning Equipment Production Tools

t | Detectable Utensils | Handling & Storage Containers

Instrumentation

