

CERTIFICATE OF ANALYSIS

Material Description: Buffer Solution pH 7 Colour Coded (Yellow)

CAS Number: Mixture

Batch Number: 883027

| Supplied As: | Product Code | Product Description |
|--------------|--------------|--|
| | PEB304-07 | Buffer Solution pH 07 Colour Coded (Yellow) 2.5L. |
| | PEB404-07 | Buffer Solution pH 07 Colour Coded (Yellow) 1L. |
| | PEB504-07 | Buffer Solution pH 07 Colour Coded (Yellow) 500ml. |

Klipspringer Ltd confirm that this material has been manufactured and tested to the following specification. Full audit trail of raw material batches and other manufacturing details can be provided on request.

| Specification | | | | | | Test Results | |
|---------------|-------|---------------|---------------|---------------|-------|--------------|--------|
| Parameter | Units | Minimum value | Maximum value | Typical value | Notes | Result | Status |
| pH @ 20°C | pH | 6.99 | 7.01 | 7.00 | | 7.002 | Pass |

| Additional Information | | | | | |
|------------------------|-------|---------------|---------------|---------------|-----------|
| Parameter | Units | Minimum value | Maximum value | Typical value | Notes |
| pH @ 10°C | pH | | | 7.05 | |
| pH @ 15°C | pH | | | 7.02 | |
| pH @ 25°C | pH | | | 6.99 | |
| pH @ 30°C | pH | | | 6.98 | |
| pH @ 37°C | pH | | | 6.97 | |
| pH @ 40°C | pH | | | 6.97 | |
| pH @ 45°C | pH | | | 6.96 | |
| pH @ 50°C | pH | | | 6.96 | |
| NIST SRM 185i | | | | | Traceable |
| NIST SRM 186-I-g | | | | | Traceable |
| NIST SRM 186-II-g | | | | | Traceable |

Doc. tracking ref: 2627/CA/883027/001
Issued by: Radek Tameczka
Position: Laboratory Manager
Issue Date: 7th July 2023

Certificate generated electronically and is valid without a signature.

Klipspringer Ltd. Registered in England and Wales: Klipspringer Ltd, Rynor House, Farthing Road, Ipswich, IP1 5AP. Registration number: 07676073.